

Ankle-brachial index and simultaneous dual arm measurement function for accurate office blood pressure measurement.

[Instruction Manual](#)



WatchBP Office ABI is a professional automatic office blood pressure measurement device which can determine the ankle brachial index (ABI) and inter arm difference (IAD) fast and accurately. The ABI is a measure for the assessment of peripheral arterial disease. The WatchBP Office ABI can perform automated dual-arm measurements for the most reliable IAD. In addition, the WatchBP office ABI can detect atrial fibrillation^{1,2} with high reliability. ABI, IAD and Afib are all important prognostic indicators for cardiovascular risk which make the WatchBP office ABI an important tool in the screening for cardiovascular risks.

1 Joseph Wiesel, Lorenzo Fitzig, Yehuda Herschman and Frank C. Messineo. Detection of Atrial Fibrillation Using a Modified Microlife Blood Pressure Monitor. *American Journal of Hypertension* 2009, 22, 848-852.

2 G S Stergiou, N Karpettas, A Protogerou, E G Nasothimiou and M Kyriakidis. Diagnostic accuracy of a home blood pressure monitor to detect atrial fibrillation. *Journal of Human Hypertension* 2009, 23, 654-658.

Table of Contents

Product description

Name of parts and display	6-7
WatchBP ABI, components and accessories	8-9

Before using WatchBP Office ABI

Selecting and fitting the correct cuff	10-11
Application of the ankle cuff	12

Three operation modes

«ABI» Mode	13
«ROUTINE» Mode	14
«SCREEN» Mode	15-16
Inter-arm difference (IAD)	16

Taking measurements using WatchBP Office ABI

«SCREEN» Mode	17-18
«ROUTINE» Mode	19-20
Skipping the countdown time	21
Taking less than three measurements	21
Setting measurement intervals	22
«ABI» Mode	23-24

Special Functions

Hiding measurement values	25
Mean arterial pressure and pulse pressure	26
Atrial fibrillation detection	27-28
ABI (Ankle Brachial Index)	29

Viewing and transferring Measurements


Viewing measurements in memory	30-31
Viewing MAP and PP	31
Transferring measurements	32-33

Appendix


Rechargeable battery and power adaptor	34
Troubleshooting	35
Error messages	36-37
Safety, care, accuracy test and disposal	38-39
Technical specifications	40

Product description

Name of parts




Display




WatchBP Office ABI, components and accessories



WatchBP Office ABI
Blood Pressure Device




AC Adaptor X1
(Input: 100-240V~50/60Hz 0.48A
Output: +7.5V 2A)



Serial Port Connecting Cable



Cuff for upper arm
M size (22cm~32cm) X2
L size (32cm~42cm) X2




Cuff for ankle
M size (22cm~32cm) X1



Instruction Manual X 1
Quick Start Guide X 1

Selective Cuffs for upper arm and ankle


For upper arm



M (Medium size, standard delivered)

22 - 32 cm (8.7 - 12.6 inches)


With air tube 130 cm



L (Large size, standard delivered)

32-42 cm (12.6-16.5 inches)

With air tube 130 cm




L-XL (Large to extra large size, to be ordered)

32-52 cm (12.6-20.5 inches)

With air tube 130 cm


For ankle



M (Medium size, standard delivered)

22 - 32 cm (8.7 - 12.6 inches)

With air tube 200 cm



L (Medium size, to be ordered)

32 - 42 cm (12.6 - 16.5 inches)

With air tube 200 cm

※ Please contact Microlife or its distributor to purchase L-XL size or other size cuffs.

Before using WatchBP Office ABI

Selecting the correct cuff

Two different sizes of cuffs for upper arms are provided with the WatchBP Office ABI: Medium and Large. Use the cuff marker to select the cuff size that best matches the circumference of the patient's upper arm.



M (Medium size)

22 - 32 cm (8.7 - 12.6 inches)

With air tube 130 cm




L (Large size)

32-42 cm (12.6 - 16.5 inches)


With air tube 130 cm

Connect the cuff to the device by inserting the cuff connector into the cuff connector socket.



Fitting the cuff properly


- 1) Place the cuff over the left (or right) upper arm so that the air tube and artery mark arrow point toward the lower arm.
- 2) Lay the cuff on the arm. Make sure that the lower edge of the cuff lies approximately 2 to 3 cm ($\frac{3}{4}$ to 1 inch) above the elbow.




- 3) Wrap and tighten the cuff around the arm.
- 4) Leave a little free space between the arm of the patient and the cuff. Two fingers should fit between the arm and the cuff. Clothing must not constrict the arm. Remove all clothing covering or constricting the measurement arm.



- 5) It may result in incorrect blood pressure readings if the cuffs are not fitted properly. Use a different size cuff if the range Index at the end of the cuff does not fall into the range specified by the range stripes.
- 6) Apply the same steps to the other arm if a double arm measurement is required.




✱ *The range index of cuff should fall into this range.*




Application of the ankle cuff (in ABI mode)

- 1) The patient has to lie down in supine position.
- 2) Place the ankle cuff on the leg. Make sure the edge of the ankle cuff lies approximately 2 to 3cm ($\frac{3}{4}$ to 1 inch) above the ankle and notice that the artery mark is on the posterior tibial artery.




- 3) Wrap and tighten the cuff around the leg.
- 4) Leave a little free space between the leg of the patient and the cuff. Two fingers should fit between the leg and the cuff.



Three operation modes


Select an operation mode

The WatchBP Office ABI device includes three operation modes: «**ABI**», «**ROUTINE**», and «**SCREEN**» (Inter-arm difference). Use the Mode Switch to select the desired mode.



«ABI» Mode


Select «**ABI**» Mode for Ankle-Brachial measurement. Select the arm with the higher blood pressure value as determined with the «**SCREEN**» Mode



Three operation modes (cont.)

«ROUTINE» Mode


Select «**ROUTINE**» Mode to perform automated duplicate measurements on the preferred arm for prompt and accurate office measurements.



ABI ROUTINE SCREEN

Automated triple measurements


In «**ROUTINE**» Mode, the WatchBP Office ABI device automatically takes three consecutive measurements at 15 second intervals on default. The readings of these three measurements are averaged to produce the routine visit blood pressure measurement.



※ *The user can manually set measurement intervals of 15, 30, 45 or 60 seconds in ROUTINE Mode. (Please refer to special functions section page 19 "Setting measurement intervals")*


«SCREEN» Mode

Select «**SCREEN**» Mode to complete fully-automated triple measurements on both arms according to recommended ESH/AHA blood pressure measurement protocols for a patient's first office visit.



Simultaneous dual-arm measurements


In «**SCREEN**» Mode, the WatchBP Office ABI device measures the patient's blood pressure on both arms simultaneously, which helps to determine the arm with the highest blood pressure value and reveals other potential cardiovascular risks.



Three operation modes (cont.)

Automated triple measurements



In «**SCREEN**» Mode, the WatchBP Office ABI device automatically takes three consecutive measurements on both arms at fixed one minute* intervals. The results of these three measurements are then averaged to conclude the blood pressure measurement.



- * The measuring intervals are fixed at one minute in **SCREEN** mode.
- * The arm with the higher blood pressure value should be taken for ABI measurement or future blood pressure measurements.

Determine the proper arm and inter-arm difference


If the difference of blood pressure readings between two arms is more than 20mmHg for Systolic or 10mmHg for Diastolic at all three measurements, the device indicates the arm with the higher blood pressure value by displaying L or R and the “IAD” icon. The “IAD” icon and readings of the higher arm flash to indicate that more attention is needed.




Taking measurements using WatchBP Office ABI

«SCREEN» Mode

- 1) **Turn on the power** – Turn on the device by switching the I/O switch at the back of the device to the ON position.




- 2) **Apply the cuff** – Properly fit one cuff to each of the patient's arms.



※ Additional visual instruction can be found on the cuff

- 3) **Switch to «SCREEN» Mode** – Slide the Mode Switch to «SCREEN» Mode.




- 4) **Inflation by fuzzy logic** – The device will adjust inflation pressure by fuzzy logic technology.

Taking measurements using WatchBP Office ABI (cont.)

- 5) **Start measurements** – Press the Start/Stop Button to start the measurement sequence.

✧ A 60-second countdown will proceed the first measurement.




- 6) **Three consecutive measurements** – The device will take three consecutive measurements with one minute intervals between measurements.



- 7) **The average value of the measurements** – The average value will be displayed after the measurements.

If one of the three measurements failed, a fourth measurement will be conducted.




✧ If the third measurements are cancelled by pressing the Start/Stop after the second measurement, the averages are calculated and saved to the memory as well.


✧ Please refer to special functions section page 18 "Taking less than three measurements"

«ROUTINE» Mode

- 1) **Turn on the power** – Turn on the device by switching the I/O switch at the back of the device to the ON position.




- 2) **Apply the cuff** – Properly fit one cuff to the preferred measurement arm.



✧ *Additional visual instruction can be found on the cuff*

- 3) **Switch to «ROUTINE» Mode** – Slide the Mode Switch to «ROUTINE» Mode.




- 4) **Select the inflation cuff** – Set the device to the preferred measurement side (left or right). Press and hold the M Button for 3 seconds.



**Press and hold
for 3 seconds...**

Taking measurements using WatchBP Office ABI (cont.)

- 5) Press the M Button to switch between left(L) and right(R). Press the Start/Stop Button to save the setting.



- 6) **Start measurements** – Press the Start/Stop Button to start the measurements.



- 7) **Three consecutive measurements** – The device will take three consecutive measurements with 15 second intervals in between on default.



- 8) **Inflation by fuzzy logic** – The device will adjust inflation pressure by fuzzy logic technology.
- 9) **The average value of the measurements** – The average values will be displayed when the measurements are completed and will be saved to the memory automatically.

※ *The user can manually set the measurement intervals of 15, 30, 45 or 60 seconds in ROUTINE Mode. (Please refer to special functions section page 19 "Setting up measurement intervals")*

Skipping the countdown time

The 60 seconds countdown before measurement in «**SCREEN**» Mode and 15 seconds in «**ROUTINE**» Mode can be skipped by pushing the Start/Stop Button. While the Start/Stop Button is pushed, the device will immediately begin the next measurement.



Skip the countdown time and begin measurement.

※ *The device can be set into Stand-by Mode by pressing the Start/Stop Button after the completion of measurements. The device will automatically switch to Stand-by Mode if left unattended for five minutes.*



Stand by mode

Taking less than three measurements

The measurement sequence can be stopped at anytime by pushing the Start/Stop Button in «**SCREEN**» or «**ROUTINE**» Mode. The device enters stand-by and the remaining measurements are cancelled. Average of the measurements is saved and can be viewed by pushing the M Button.



Cancel remaining measurements at anytime during the measurement sequence.

Taking measurements using WatchBP Office ABI (cont.)


Setting measuring intervals in «ROUTINE» Mode

- 1) Pressing and holding the Start/Stop Button for 3 seconds.



**Press and hold
for 3 seconds...**


- 2) Press the M Button to adjust the measurement interval, then press the Start/Stop Button to confirm, the device will go back to stand-by Mode.




** The default measuring interval is 15 seconds. The interval can be set as 15, 30, 45 or 60 seconds.*

«ABI» Mode


- 1) **Turn on the power** – Turn on the device by switching the I/O switch at the back of the device into the ON position.



- 2) **Switch to «ABI» Mode** – Slide the Mode Switch to «ABI» Mode. The supine position reminder will display.




- 3) The patient has to lie down in supine position.
- 4) **Apply the cuff*** – Properly fit the arm cuff to the upper arm and the ankle cuff to the leg of the preferred measurement side of the body. The preferred side is determined in «**SCREEN**» Mode.
- 5) Make sure the upper arm cuff is connected to the left cuff socket and the ankle cuff is connected to the right cuff socket.



* Additional visual instruction can be found on the cuff.

Taking measurements using WatchBP Office ABI (cont.)


- 6) **Start measurements** – Press the Start/Stop Button to start the measurement.



- 7) **Inflation by fuzzy logic** – The device will adjust inflation pressure automatically by fuzzy logic technology.

- ∗ There is only one set of memory capacity in the device. Only the last measurements will be saved in the memory.
- ∗ The memory will be deleted when the power is turned off.


- 8) **Display of the measurement values**– The measurement values will be displayed and the Ankle-brachial index is automatically calculated when the measurement is completed. The Ankle-brachial index and the blood pressure value display first, the brachial index value is replaced by the pulse value for 3 seconds. All the values are saved automatically.



Special Functions

Hiding measurement values

The WatchBP Office ABI device features a Hide function in order to prevent unnecessary elevated blood pressure in patients due to nervousness triggered by visible blood pressure values.





※ This feature is available in both *SCREEN* mode and *ROUTINE* mode.

※ When the «**Hide**» function is ON, only the average values are displayed after the consecutive measurements and the Hide icon is displayed.


- 1) **Activate the «Hide» Function** – Press M button; L or R flashes; keep pressing and holding the M Button for 7 more seconds until On or Off flashes.



- 2) **Select ON or OFF** – Press the M Button again to turn the “Hide” function ON or OFF.



- 3) **Confirm** – Press Start/Stop to confirm the setting.




※ The default setting of “Hide” is set at “ON” in both «*SCREEN*» Mode and «*ROUTINE*».

Special Function (cont.)

MAP (Mean Arterial Pressure)


The WatchBP Office ABI device measures the true mean arterial pressure (MAP) of the patient. Each measurement includes a single MAP value. The average measurement will display the average MAP value. When viewing in Memory mode, the MAP value will be displayed with the systolic and diastolic pressure once every 5 seconds.



∗ The Mean Arterial Pressure (MAP) in this device is determined from the maximum peak of the oscillometric envelope curve.

PP (Pulse Pressure)


The WatchBP Office ABI device provides the pulse pressure (PP) of the patient: pulse pressure = systolic - diastolic pressure. Each measurement includes the calculation of a single PP value. The average measurement will display the average PP value. When viewing in Memory mode, the PP value will be displayed with the systolic/diastolic pressure once every 5 seconds.



Appearance of the atrial fibrillation indicator for early detection

The WatchBP Office ABI is designed to detect asymptomatic atrial fibrillation during blood pressure measurements in the «**SCREEN**» and «**ROUTINE**» Mode. If two out of three measurements detect atrial fibrillation, the Afib icon is displayed. This device is able to detect atrial fibrillation with high accuracy: a sensitivity of 97% and a specificity of 89%. ✱ If the atrial fibrillation icon is displayed after the measurement, it is recommended to visit a doctor.

- ✱ Joseph Wiesel, Lorenzo Fitzig, Yehuda Herschman and Frank C. Messineo. *Detection of Atrial Fibrillation Using a Modified Microlife Blood Pressure Monitor. American Journal of Hypertension 2009; 22, 848-852.*
- ✱ G S Stergiou, N Karpettas, A Protogerou, E G Nasothimiou and M Kyriakidis. *Diagnostic accuracy of a home blood pressure monitor to detect atrial fibrillation Home monitor for atrial fibrillation. Journal of Human Hypertension 2009, 23, 654-658.*



- ✱ This device detects atrial fibrillation, a major cause of stroke. Not all risk factors for stroke, including atrial flutter, may be detected by this device.
- ✱ This device may not detect atrial fibrillation in people with pacemakers or defibrillators.

About Atrial Fibrillation

Atrial fibrillation is a common heart rhythm problem. It affects more than 2 million people in North America. It is more common in older age and it is found in 18% individuals aged 85 years and older. It is a common cause of major strokes. About 15% of all strokes are caused by atrial fibrillation.

The elderly, or those with high blood pressure, diabetes or heart disease are more likely to get a stroke if they have atrial fibrillation.

Atrial fibrillation is a rhythm problem that can last from a few minutes, to days or weeks and even years. Atrial fibrillation can cause blood clots in the upper chambers of the heart (the atria). These clots can break off and flow to the brain causing a stroke.

The use of blood thinners, such as warfarin, can lower the risk of a stroke in patients with atrial fibrillation.

A doctor can confirm the presence of Afib by using an ECG. Sometimes Afib is present incidentally. Therefore, a doctor may not see it on regularly scheduled visits.


One method of detecting Afib is by mean of palpations. This method is not very reliable. Failure of detecting Afib may lead to the occurrence of a stoke in the end; whereas early detection may lead to early treatment that can significantly reduce the chances of a stroke.

ABI (Ankle Brachial Index)

The systolic blood pressure values of a person's arm and leg need to be measured in order to determine ABI (Ankle Brachial Index).

The ankle-brachial index (ABI) is then calculated using the quotient of the systolic pressure from the leg measurement and the systolic pressure from the arm measurement.


A low ankle-brachial index (ABI) is indicative of systemic vascular disease, and should place a patient in the high-risk category.



Viewing and transferring measurement readings

«ROUTINE» and «SCREEN» Mode


- 1) **Switch the mode switch** – The M Button can be pressed in either the «ROUTINE» or «SCREEN» mode to retrieve the last set of readings from the measurements.



- 2) **Viewing the average of all measurements** – At first an “A” will be displayed, followed by an average of all measurements taken in the last sequence of measurements.




- 3) **Viewing MAP and PP values** – Ten seconds after the display of the average blood pressure, the device will display values for Pulse Pressure (PP) and Mean Arterial Pressure (MAP).




- 4) **Viewing individual measurements** – Press the M Button once again to view individual measurements. The display will flash a number “3” indicating the third measurement.



- 5) Values of the third measurement (including PP and MAP values) will be displayed in the order listed in step 3.




- 6) Press the M Button to retrieve values for second and first measurements from the last completed set of measurements.



«ABI» Mode

- 1) **Switch the mode switch** –Select the slide switch to «ABI» Mode then press M button to retrieve the last sets readings of the measurements.
- 2) **Viewing the Ankle-brachial index (ABI)**
At first the INDEX icon, the ABI and the blood pressure values will be displayed.
- 3) **Viewing MAP and PP values** –
Ten seconds after the display of the ABI, the values of Pulse Pressure (PP) and Mean Arterial Pressure (MAP) will be displayed.



Viewing and transferring Measurements (cont.)

Transferring measurements

Installation of the software program


- 1) Put the CD in the CD-ROM drive of your computer or click on «**setup.exe**» in the CD's directory alternatively.
- 2) Follow the instructions provided in the installation window on the computer screen.
- 3) When installation is finished, be sure to restart the computer before you start working with the program.



❖ *System Requirements:*
550MHz CPU, 256MB Memory,
1024x768 pixel resolution,
256 color, CD-ROM drive, 1 free
USB port, 40MB free hard disk
space, Microsoft Windows 7/
8/ 10.

Transferring data to the computer

- 1) Start the software program and connect the device to the computer using the cable supplied.
- 2) A successful connection is displayed by «**Connected**» on the computer screen.
- 3) Enter identity number, Patient Name and Surname to create a new record.



Software commands		Refer to the software user manual for detailed information and instructions.
Perform a measurement		Click «Measure»
Stop a measurement		Click «Stop»
Store data		Click «Save»
Close		Click «Close»

✧ *Please read the software manual to know more details of the software operation commands.*

Appendix

Rechargeable Battery

WatchBP Office ABI equips with a built-in, rechargeable Ni-MH battery pack which delivers up to 400~500 measurement cycles. The batteries can be recharged while the power adaptor is plugged in. It doesn't affect the measurement.




- ❖ *First time use, charge the batteries until the recharge indicator turns to green*
- ❖ *An orange recharge indicator means recharge in progress.*
- ❖ *A green recharge indicator means recharge is completed.*

Using a power adaptor

Please only use the Microlife adaptor supplied with the WatchBP Office ABI to recharge the device.

- 1) Plug the adapter cable into the Power Plug of the WatchBP Office ABI device.
- 2) Plug the adaptor plug into the wall socket. When the power adaptor is connected, no battery power is consumed.



Troubleshooting


Problem	Possible cause	How to make correction
No power(No LCD display)	Power supply is not properly plugged in	Plug power supply into wall socket.
	Battery is fully discharged	Recharge the rechargeable battery by plugging in the power supply.
Cuff does not inflate properly	Loose connection of the tube	Make sure the tube of the cuff is securely connected to the device.
	Leakage of the tube or bladder	Check for cracks on the tube or the bladder. Please contact Microlife customer service for this issue.
No result displayed between measurements	The Hide function is activated	Disable “hide” function, or use «ROUTINE» mode to measure the blood pressure.

Error messages

If an error occurs during measurement, the measurement is interrupted and an error message «Err» is displayed.



- Please contact your local Microlife service center if the error persists.
- If you think the results are unusual, please read through the information in this instruction manual carefully.



Error	Description	Potential cause and remedy
«Err 1»	Signal too weak	The pulse signals on the cuff are too weak. Re-position the cuff and repeat the measurement.
«Err 2»	Error signal	During the measurement, error signals were detected by the cuff, caused for instance by movement or muscle tension. Repeat the measurement, keeping your arm still.

«Err 3»	No pressure in the cuff	An adequate pressure cannot be generated in the cuff. A leak may have occurred. Replace the batteries if necessary. Repeat the measurement.
«Err 5»	Abnormal result	The measuring signals are inaccurate and no result can therefore be displayed. Read through the checklist for performing reliable measurements and then repeat the measurement.

«HI»	Pulse or cuff pressure too high	The pressure in the cuff is too high (over 300 mmHg) OR the pulse is too high (over 200 beats per minute). Relax for 5 minutes and repeat the measurement.
«LO»	Pulse too low	The pulse is too low (less than 40 beats per minute). Repeat the measurement.

Safety, care, accuracy test and disposal

Safety and protection

This device may be used only for the purpose described in this booklet. The device comprises of sensitive components and must be treated with caution. The manufacturer cannot be held liable for damage caused by incorrect application.



- Ensure that children do not use the device unsupervised; some parts are small enough to be swallowed.
- Only activate the pump when cuff is installed.
- Do not use the device if you think it is damaged or if anything appears unusual.
- Read the further safety instructions in the individual sections of the instruction manual.
- Do not connect the device to a computer until prompted to do so by the computer software.

Observe the storage and operating conditions described in the “Technical specifications” section of this manual.



Protect the device from water and moisture



Protect the device from direct sunlight



Protect the device from extreme heat and cold



Avoid proximity to electromagnetic fields, such as those produced by mobile phones



Never open device




Protect device from impact and drops

Device care

Use a soft cloth with one of the following recommended cleaning solutions to wipe the exterior of the device:

- Ethyl or isopropyl alcohol (70% solution)
- Hydrogen peroxide 7.5% solution
- Sodium hypochlorite solution (5.25-6.15% household bleach diluted 1:500 provides >100 ppm available chlorine)




Accuracy test

We recommend the WatchBP Office ABI device be tested for accuracy every 2 years or after mechanical impact (e.g. being dropped). Please contact Microlife to arrange for an accuracy test.

Cleaning the cuff

Take out the bladder. Fold and place the cuff cover inside a washing bag. Wash cuff cover with warm water and a mild detergent in washing machine. Air dry the cuff. DO NOT iron the cuff cover.



Do not iron the cuff!



Disposal

Batteries and electronic instruments must be disposed of in accordance with the locally applicable regulations, and not as domestic waste.

Technical specifications

**Operation temperature/
humidity:**

- 10 to 40 °C (50 to 104 °F)

**Storage temperature/
humidity:**

- -20 to 55 °C (-4 to 131 °F)
- 15 - 90 % relative maximum humidity

Weight:

- 1100 g (including rechargeable battery pack)

Dimensions:

- 200 x 125 x 90 mm

Measuring method:

- Oscillometric, corresponding to Korotkoff

Measurement range:

- 30 - 280 mmHg – blood pressure
- 40 - 200 beats per minute – pulse

Cuff pressure display:

- Range: 0 - 299 mmHg
- Resolution: 1 mmHg
- Static accuracy: pressure within ± 3 mmHg
- Pulse accuracy: ± 5 % of the readout value

Voltage source:

- Rechargeable battery pack; 4.8V C3500 mAh
- Mains adapter DC 7.5V, 2 A

Reference to Standards:

- Device corresponds to the requirements of the standard for non-invasive blood pressure monitor.
EN 1060-1
EN 1060-3
EN 1060-4
IEC 60601-1
IEC 60601-1-2
- Device fulfills the stipulations of the standard IEC 60601-1-2.



Type BF applied
part

Microlife reserves the right to alter technical specifications without prior written notice.

Guarantee card

This device is covered by a two-year guarantee and accessories are covered by a one-year guarantee from the date of purchase. This guarantee is valid only on presentation of the guarantee card completed by the owner confirming date of purchase or purchase receipt.


Name: _____

Address: _____

Date: _____

Telephone: _____


E-mail: _____



Product: WatchBP Office ABI

Product Number: TWIN200 ABI

Date:



**Europe / Middle-East /
Africa**

Microlife WatchBP AG
Esenstrasse 139
9443 Widnau, Switzerland
Tel +41 71 727 7000
Fax +41 71 727 7011
Email: watchbp@microlife.ch
www.watchbp.com

Asia

Microlife Corporation
9F, 431, RuiGuang Road, NeiHu
Taipei, 11492, Taiwan, R.O.C.
Tel +886 2 8797 1288
Fax+886 2 8797 1283
Email:
watchbp@microlife.com.tw
www.watchbp.com

United States, Canada

Microlife Medical Home
Solutions, Inc.
2801 Youngfield St., Suite 241
Golden, CO 80401, USA
Tel. +1 303 274 2277
Fax +1 303 274 2244
Email: watchbp@mimhs.com
www.watchbp.com

**Central / South
America**

Microlife USA, Inc.
1617 Gulf To Bay Blvd., 2nd Floor
Clearwater, FL 33755, USA
Tel +1 727 442 5353
Fax +1 727 442 5377
Email: msa@microlifeusa.com
www.watchbp.com



Read the instruction manual carefully before using this device, especially the safety instructions, and keep the instruction manual for future use.

Caution: Federal law restricts this device to sale by or on the order of a physician.

IB WatchBP Office ABI EN 2520