



# Declaration of Conformity

The manufacturer, A&D Company, Limited, is exclusively responsible for the declaration of conformity.

**Manufacturer 's Name and Address:**

A&D Company, Limited  
1-243 Asahi, Kitamoto-shi, Saitama-ken. 364-8585 Japan

**European Representative:**

A&D INSTRUMENTS LIMITED  
Unit24/26 Blacklands Way  
Abingdon Business Park  
Abingdon, Oxon  
OX14 1DY United Kingdom

**Product Category:**

Blood pressure monitor

**Model Number:**

TM-2441

**Classification/Rule:**

II a/ Rule10

- ✓ **The product complies with the requirements of the Directive 93/42/EEC Annex I and the certified quality system according to Annex II excluding (4).**

**Standards:**

EN 60601-1:2006/A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010  
EN 60601-1-11:2010, EN 62304:2006, EN 62366:2008,  
IEC 80601-2-30:2009/A1:2013, EN ISO 10993-1:2009, ISO 81060-2:2013  
EN ISO 14971:2012, EN 1041:2008/A1:2013, EN ISO 15223-1:2016

**Notified Body:**

TÜV SÜD Product Service GmbH (0123)  
Ridlerstraße 65 80339 MÜNCHEN, GERMANY

**Certificate number: G1 085349 0008 Rev. 00**

This declaration is valid until the expiration date of the MDD annex II certificate by TÜV SÜD Product Service GmbH.

✓ **The product complies with the requirements of the Directive 2014/53/EU.**

**Standards:**

EN 300 328 V2.1.1


**Notified Body: Not Applied**

✓ **The product complies with the requirements of the Directive 2011/65/EU.**

**Notified Body: Not Applied**

**Place and date:** Saitama, Japan, 14 Mar. 2019

**File number:** 19031406



---

**Minoru Saitou**

**Chief of Quality Control Office**

**A&D Company, Limited**